



### CALIBRATION CERTIFICATE

<b>Customer:</b>	<b>Address:</b>		
<b>Description:</b> Optical Time Domain Reflectometer (OTDR)			
<b>Serial no.:</b> 922031	<b>Model no.:</b> 720C-Q1-EA-EI		
<b>Calibration location:</b> 1/F, Block C, Funing High-Tech Park, No. 71, Xintian Avenue, Fuyong, Shenzhen, China, 518103			<b>Calibration date:</b> 2017-09-12

As found					
<input type="checkbox"/> Initial calibration	<input checked="" type="checkbox"/> Within specifications (i)	<input type="checkbox"/> Within specifications* (ii)	<input type="checkbox"/> Outside specifications* (iii)	<input type="checkbox"/> Outside specifications (iv)	<input type="checkbox"/> Defective operation
*See results page for details on "As found" status					
Comments on unit status:					
Action taken					
<input checked="" type="checkbox"/> No adjustment was made		<input type="checkbox"/> Adjustments were made		<input type="checkbox"/> Repair was performed	
As left					
<input checked="" type="checkbox"/> Within specifications			<input type="checkbox"/> Outside specifications		
Calibration conditions					
Temperature: 23 °C ± 2 °C					
Standard(s) used to establish traceability					
Description	Inventory number	Calibration date	Calibration validity		
Standard(s) used for "As found" measurements					
METAS Length Reference Fiber MM	LRF2015.08	2015-06-25	Indefinite		
METAS Length Reference Fiber SM	LRF2015.14	2015-09-24	Indefinite		
Standard(s) used for "As left" measurements					
METAS Length Reference Fiber MM	LRF2015.08	2015-06-25	Indefinite		
METAS Length Reference Fiber SM	LRF2015.14	2015-09-24	Indefinite		

Remarks:

Optical ports are always cleaned before calibration.  
EXFO certifies that the unit has been calibrated using standards traceable to a national metrology institute (NIST, NPL, NRC, METAS or other), natural physical constants or using ratio measurements. NIST is the National Institute of Standards and Technology in the USA, NPL is the National Physical Laboratory in the UK, NRC is the National Research Council in Canada and METAS is the Swiss Federal Office of Metrology. All uncertainties are reported with a level of confidence of 95 %. Calibration is based on the ISO/IEC 17025 standard.  
The certificate shall not be reproduced, except in full, without the written approval of EXFO.

**EXFO**  
Serial No.: 922031  
Cal./test date: 2017-09-12 | Cal./test due:

\_\_\_\_\_  
**Xiaolian Lan**  
Calibration operator

\_\_\_\_\_  
2017-09-12  
Date

<b>Dynamic range test (Single Mode)</b>			
Procedure: IETA-00366 (pulse: 10 $\mu$ s, range: 240 km, time: 45 s)			
Nominal wavelength (nm)	Measured dynamic range (dB)	Test limit (dB)	Verification
<b>As found</b>			
1310	37.3	35.5	Pass
1550	35.4	33.9	Pass
<b>As left</b>			
1310	37.3	35.5	Pass
1550	35.4	33.9	Pass

<b>Attenuation dead zone test (Single Mode)</b>			
Procedure: IETA-00366 (pulse: 10 ns, range: 0.65 km, reflectance: -45 dB)			
Nominal wavelength (nm)	Measured attenuation dead zone (m)	Test limit (m)	Verification
<b>As found</b>			
1310	4.17	4.99	Pass
1550	4.38	5.11	Pass
<b>As left</b>			
1310	4.17	4.99	Pass
1550	4.38	5.11	Pass

<b>Event dead zone test (Single Mode)</b>			
Procedure: IETA-00366 (pulse: 3 ns, range: 0.65 km, reflectance: -45 dB)			
Nominal wavelength (nm)	Measured event dead zone (m)	Test limit (m)	Verification
<b>As found</b>			
1310	0.67	0.78	Pass
1550	0.67	0.81	Pass
<b>As left</b>			
1310	0.67	0.78	Pass
1550	0.67	0.81	Pass



<b>Distance calibration (Single Mode)</b>							
Procedure: IETA-00366 (pulse: 30 ns, range: 3.5 km, time: 15 s)							
Nominal wavelength (nm)	Measured length (m)	Reference length (m)	Deviation (m)	Uncertainty (m)	Specification (m)	Conformance limit (m)	Verification*
<b>As found</b>							
1310	2349.03	2348.98	0.05	0.06	0.97	0.92	i
1550	2349.83	2349.98	-0.15	0.28	0.97	0.74	i
<b>As left</b>							
1310	2349.03	2348.98	0.05	0.06	0.97	0.92	i
1550	2349.83	2349.98	-0.15	0.28	0.97	0.74	i

<b>Linearity test (Single Mode)</b>		
Procedure: IETA-00366 (wavelength: 1310 nm, pulse: 10 $\mu$ s, range: 21.8 dB, time: 45 s)		
Measured non linearity (dB)	Test limit (dB)	Verification
<b>As found</b>		
0.01	0.03	Pass
<b>As left</b>		
0.01	0.03	Pass

<b>Dynamic range test (Multi Mode)</b>			
Procedure: IETA-00366 (pulse: 1 $\mu$ s, range: 5 km, time: 45 s)			
Nominal wavelength (nm)	Measured dynamic range (dB)	Test limit (dB)	Verification
<b>As found</b>			
850	26.2	24.8	Pass
1300	28.4	26.7	Pass
<b>As left</b>			
850	26.2	24.8	Pass
1300	28.4	26.7	Pass

<b>Attenuation dead zone test (Multi Mode)</b>			
Procedure: IETA-00366 (pulse: 10 ns, range: 0.65 km, reflectance: -35 dB)			
Nominal wavelength (nm)	Measured attenuation dead zone (m)	Test limit (m)	Verification
<b>As found</b>			
850	2.77	3.36	Pass
1300	3.21	3.77	Pass
<b>As left</b>			
850	2.77	3.36	Pass
1300	3.21	3.77	Pass

<b>Event dead zone test (Multi Mode)</b>			
Procedure: IETA-00366 (pulse: 3 ns, range: 0.65 km, reflectance: -35 dB)			
Nominal wavelength (nm)	Measured event dead zone (m)	Test limit (m)	Verification
<b>As found</b>			
850	0.47	0.55	Pass
1300	0.43	0.52	Pass
<b>As left</b>			
850	0.47	0.55	Pass
1300	0.43	0.52	Pass

Distance calibration (Multi Mode)							
Procedure: IETA-00366 (pulse: 30 ns, range: 3.5 km, time: 15 s)							
Nominal wavelength (nm)	Measured length (m)	Reference length (m)	Deviation (m)	Uncertainty (m)	Specification (m)	Conformance limit (m)	Verification*
<b>As found</b>							
850	114.70	114.69	0.01	0.11	0.91	0.83	i
1300	114.22	114.24	-0.02	0.08	0.91	0.85	i
<b>As left</b>							
850	114.70	114.69	0.01	0.11	0.91	0.83	i
1300	114.22	114.24	-0.02	0.08	0.91	0.85	i

Linearity test (Multi Mode)		
Procedure: IETA-00366 (wavelength: 850 nm, pulse: 1 µs, range: 12.9 dB, time: 45 s)		
Measured non linearity (dB)	Test limit (dB)	Verification
<b>As found</b>		
-0.01	0.03	Pass
<b>As left</b>		
-0.01	0.03	Pass

OTDR Self-Test
Procedure: IETA-00366
Status (Pass/Fail)
<b>As found</b>
Pass
<b>As left</b>
Pass

\* Verification status legend:

- i) Within specifications;
- ii) Within specifications: All measured results are within specifications limits. In conformance with ISO/IEC 17025, full compliance cannot be stated because of measurement uncertainties. Nevertheless, results indicate that the instrument is likely to perform according to specifications;
- iii) Outside specifications: Some measured results are outside specification limits. Nevertheless, as per ISO/IEC 17025, non compliance cannot be stated because of measurement uncertainties;
- iv) Outside specifications.

Unless otherwise stated, 100 % of shipped units have all "As left" results in case i.